

MRI (Depiction of Intramedullary Hyperintensity at Eight Cervical Disc Levels, T2W, 1.5-T or 3-T)

EVIDENCE

The Depiction of Intramedullary Hyperintensity at Eight Cervical Disc Levels Using T2W 1.5 or 3-T MRI has a category B recommendation in accordance with COSMIN criteria (1-3).

RELIABILITY

Property	Result Summary	Overall Rating	Quality of Evidence	Reference
Reliability	Inter-observer reliability: Kendall's W range 0.72–0.78; total sample size: 79	Indeterminate	Very low	(45)

VALIDITY

- No formal assessment found

RESPONSIVENESS

- No formal assessment found

INTERPRETABILITY

- No formal assessment found

FEASIBILITY

Tool	Time (Min)	Equipment	Training	License	Money	Ease of Administration	Overall Assessment
MRI (Depiction of Intramedullary Hyperintensity at Eight Cervical Disc Levels, T2W, 1.5-T or 3-T)	5-15	Minimal	No	No	No	Minimal barriers	Minimal barriers